

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

### SECTION 1: Identification of the substance/mixture and of the company/undertaking

- · 1.1 Product identifier
- · Trade name: Callisto Implant 60
- · 1.2 Relevant identified uses of the substance or mixture and uses advised against

No further relevant information available.

- · Application of the substance / the mixture Manufacture of dental prothesis
- · 1.3 Details of the supplier of the safety data sheet
- · Manufacturer/Supplier:

Ivoclar Vivadent Inc.

175 Pineview Drive, Amherst, N.Y. 14228

USA

Tel. +1 800 533 6825

Fax +1 716 691 2285

· Further information obtainable from:

Ivoclar Vivadent AG, FL-9494 Schaan, Liechtenstein

Regulatory Affairs

sds@ivoclarvivadent.com

• 1.4 Emergency telephone number: +423/235 33 13 (Ivoclar Vivadent AG, FL-9494 Schaan, Liechtenstein)

### SECTION 2: Hazards identification

- · 2.1 Classification of the substance or mixture
- · Classification according to Regulation (EC) No 1272/2008

Ox. Sol. 2 H272 May intensify fire; oxidiser.

- · 2.2 Label elements
- · Labelling according to Regulation (EC) No 1272/2008

The product is classified and labelled according to the CLP regulation.

· Hazard pictograms



GHS03

- · Signal word Danger
- · Hazard statements

H272 May intensify fire; oxidiser.

· Precautionary statements

P210 Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking.

*P221* Take any precaution to avoid mixing with combustibles.

*P280* Wear protective gloves/protective clothing/eye protection/face protection.

*P220 Keep/Store away from clothing/combustible materials.* 

P370+P378 In case of fire: Use for extinction: CO2, powder or water spray.

P501 Dispose of contents/container in accordance with local/regional/national/international

regulations.

· Additional information:

Alloys do not require a label, providing they do not present a hazard to human health by inhalation, ingestion or contact with skin or to the aquatic environment in the form in which they are placed on the market.

- · 2.3 Other hazards
- · Results of PBT and vPvB assessment
- · PBT: Not applicable.

(Contd. on page 2)

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

Trade name: Callisto Implant 60

· vPvB: Not applicable.

(Contd. of page 1)

### SECTION 3: Composition/information on ingredients

- · 3.2 Chemical characterisation: Mixtures
- · Description: Dental alloy

· Dangerous components:			
CAS: 7440-05-3 EINECS: 231-115-6		Self-heat. 2, H252; Ox. Sol. 2, H272	50-100%
CAS: 7440-22-4 EINECS: 231-131-3		substance with a Community workplace exposure limit	25-50%

· Additional information: For the wording of the listed risk phrases refer to section 16.

#### SECTION 4: First aid measures

- · 4.1 Description of first aid measures
- · General information: No special measures required.
- · After inhalation:

Grinding dust: Supply fresh air; consult doctor in case of complaints.

*In case of unconsciousness place patient stably in side position for transportation.* 

- · After skin contact: After contact with the molten product, cool rapidly with cold water.
- · After eye contact:

Grinding dust: Rinse opened eye for several minutes under running water. If symptoms persist, consult a doctor.

Mechanical effects only.

- · After swallowing: Seek medical treatment.
- 4.2 Most important symptoms and effects, both acute and delayed No further relevant information available.
- · 4.3 Indication of any immediate medical attention and special treatment needed

No further relevant information available.

#### **SECTION 5: Firefighting measures**

- · 5.1 Extinguishing media
- · Suitable extinguishing agents: The product is not flammable.
- · 5.2 Special hazards arising from the substance or mixture No further relevant information available.
- · 5.3 Advice for firefighters
- · Protective equipment: Do not inhale explosion gases or combustion gases.

#### SECTION 6: Accidental release measures

- 6.1 Personal precautions, protective equipment and emergency procedures Not required.
- · 6.2 Environmental precautions: No special measures required.
- 6.3 Methods and material for containment and cleaning up: Pick up mechanically.
- · 6.4 Reference to other sections

See Section 7 for information on safe handling.

See Section 8 for information on personal protection equipment.

See Section 13 for disposal information.

GB

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

Trade name: Callisto Implant 60

(Contd. of page 2)

#### **SECTION 7: Handling and storage**

· 7.1 Precautions for safe handling

Extractors are required on all machines used for thermal processing or splinter removal processes.

- · Information about fire and explosion protection: No special measures required.
- · 7.2 Conditions for safe storage, including any incompatibilities
- · Storage:
- Requirements to be met by storerooms and receptacles: No special requirements.
- · Information about storage in one common storage facility: Not required.
- · Further information about storage conditions: None.
- $\cdot$  7.3 Specific end use(s) No further relevant information available.

### SECTION 8: Exposure controls/personal protection

- · Additional information about design of technical facilities: No further data; see item 7.
- · 8.1 Control parameters
- · Ingredients with limit values that require monitoring at the workplace:

7440-22-4 silver

WEL Long-term value: 0.1 mg/m<sup>3</sup>

- · Additional information: The lists valid during the making were used as basis.
- · 8.2 Exposure controls
- · Personal protective equipment:
- General protective and hygienic measures: Usual hygienic measures for dental practice.
- · Respiratory protection: Use respiratory protective device against the effects of fumes/dust/aerosol.
- · Protection of hands:

Protective gloves should always be worn during mechanical and thermal processing.

After use of gloves apply skin-cleaning agents and skin cosmetics.

· Material of gloves

Mechanical processing:

Leather gloves

Strong material gloves

Thermal processing:

Heat protection gloves

· Penetration time of glove material

The exact break through time has to be found out by the manufacturer of the protective gloves and has to be observed.

· Eye protection:

Always wear safety goggles during mechanical processing (grinding, sawing /cutting, drilling, milling).

#### SECTION 9: Physical and chemical properties

- · 9.1 Information on basic physical and chemical properties
- · General Information
- · Appearance:

Form: Solid
Colour: White

Odour: Odourless

Odour threshold: Not determined.

· pH-value: Not applicable.

(Contd. on page 4)

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

Trade name: Callisto Implant 60

	(Contd. of pa
· Change in condition	
Melting point/Melting range:	1100–1260 °C
Boiling point/Boiling range:	Undetermined.
· Flash point:	Not applicable.
· Flammability (solid, gaseous):	Not determined.
· Self-igniting:	Product is not selfigniting.
Danger of explosion:	Product does not present an explosion hazard.
Explosion limits:	
Lower:	Not determined.
Upper:	Not determined.
· Vapour pressure:	Not applicable.
Density at 20 °C:	10.8 g/cm³
Relative density	Not determined.
· Vapour density	Not applicable.
Evaporation rate	Not applicable.
· Solubility in / Miscibility with	
water:	Insoluble.
· Partition coefficient (n-octanol/w	pater): Not determined.
· Viscosity:	
Dynamic:	Not applicable.
Kinematic:	Not applicable.
· Solvent content:	
Solids content:	100 %
· 9.2 Other information	No further relevant information available.

## SECTION 10: Stability and reactivity

- · 10.1 Reactivity No further relevant information available.
- · 10.2 Chemical stability Stable under normal handling and storage conditions.
- · Thermal decomposition / conditions to be avoided: No decomposition if used according to specifications.
- · 10.3 Possibility of hazardous reactions No dangerous reactions known.
- · 10.4 Conditions to avoid No further relevant information available.
- · 10.5 Incompatible materials: No further relevant information available.
- · 10.6 Hazardous decomposition products: None under normal conditions of storage and use.

## SECTION 11: Toxicological information

- · 11.1 Information on toxicological effects
- · Acute toxicity:
- · Primary irritant effect:
- · on the skin: No irritant effect.
- · on the eye: No irritating effect.
- · Sensitisation: No sensitising effects known.

(Contd. on page 5)

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

Trade name: Callisto Implant 60

(Contd. of page 4)

· Additional toxicological information:

Fumes or dusts generated from cutting or grinding operations may cause respiratory irritation.

## SECTION 12: Ecological information

- · 12.1 Toxicity
- · Aquatic toxicity: No further relevant information available.
- · 12.2 Persistence and degradability No further relevant information available.
- 12.3 Bioaccumulative potential No further relevant information available.
- · 12.4 Mobility in soil No further relevant information available.
- · Additional ecological information:
- · General notes: Generally not hazardous for water
- · 12.5 Results of PBT and vPvB assessment
- · PBT: Not applicable.
- · vPvB: Not applicable.
- · 12.6 Other adverse effects No further relevant information available.

## SECTION 13: Disposal considerations

- · 13.1 Waste treatment methods
- · Recommendation

Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.

· European waste catalogue

20 01 40 metals

- · Uncleaned packaging:
- · Recommendation: Disposal must be made according to official regulations.

· 14.1 UN-Number	
· ADR, ADN, IMDG, IATA	Void
· 14.2 UN proper shipping name	
· ADR, ADN, IMDG, IATA	Void
· 14.3 Transport hazard class(es)	
· ADR, ADN, IMDG, IATA	
· Class	Void
· 14.4 Packing group	
· ADR, IMDĞ, IATA	Void
· 14.5 Environmental hazards:	
· Marine pollutant:	No
· 14.6 Special precautions for user	Not applicable.
· 14.7 Transport in bulk according to Annex	II of
MARPOL73/78 and the IBC Code	Not applicable.
· Transport/Additional information:	Product is not classified as a dangerous good for transpo. (ADR, IMDG, IATA).

(Contd. on page 6)

Printing date 23.06.2015 Version number 4 Revision: 07.05.2015

Trade name: Callisto Implant 60

(Contd. of page 5)

· UN "Model Regulation":

## SECTION 15: Regulatory information

- · 15.1 Safety, health and environmental regulations/legislation specific for the substance or mixture
- · National regulations:
- $\cdot \textit{Other regulations, limitations and prohibitive regulations}$

The product is a medical device according to the Directive 93/42/EEC.

· 15.2 Chemical safety assessment: A Chemical Safety Assessment has not been carried out.

## **SECTION 16: Other information**

This information is based on our present knowledge. However, this shall not constitute a guarantee for any specific product features and shall not establish a legally valid contractual relationship.

· Relevant phrases

H252 Self-heating in large quantities; may catch fire.

H272 May intensify fire; oxidiser.

· Abbreviations and acronyms:

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

IATA: International Air Transport Association

GHS: Globally Harmonised System of Classification and Labelling of Chemicals

EINECS: European Inventory of Existing Commercial Chemical Substances

ELINCS: European List of Notified Chemical Substances

CAS: Chemical Abstracts Service (division of the American Chemical Society)

 ${\it Self-heat.}~2: {\it Self-Heating~Substances~and~Mixtures}, {\it Hazard~Category~2}$ 

Ox. Sol. 2: Oxidising Solids, Hazard Category 2

 $\cdot$  \* Data compared to the previous version altered.

- GI