ivoclar vivadeni:

*

Material Safety Data Sheet acc. to ISO 11014

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

1 Identification

- · Product identifier
- Trade name: Callisto Implant 78

• *Relevant identified uses of the substance or mixture and uses advised against* No further relevant information available.

- · Application of the substance / the mixture Manufacture of dental prothesis
- · Details of the supplier of the safety data sheet

Manufacturer/Supplier: Ivoclar Vivadent Inc. 175 Pineview Drive, Amherst, N.Y. 14228 USA Tel. +1 800 533 6825 Fax +1 716 691 2285

Ivoclar Vivadent Inc. 1-6600 Dixie Road Mississauga, Ontario L5T 2Y2 Canada Phone: +1 905 670 8499 Fax: +1 905 670 3102

Information department: Quality Assurance / Regulatory Affairs
 Emergency telephone number:
 24 Hour Emergency Assistance:

Emergency-Call USA - Infotrac: 1-800-535-5053 Emergency-Call Canada - Canutec: 1-613-996-6666

General SDS Assistance: US: 1-800-533-6825 Canada: 1-800-263-8182

2 Hazard(s) identification

· Classification of the substance or mixture

Ox. Sol. 2 H272 May intensify fire; oxidizer.

- · Label elements
- · GHS label elements

The product is classified and labeled according to the Globally Harmonized System (GHS).

· Hazard pictograms



· Signal word Danger

- · Hazard statements
- May intensify fire; oxidizer.
- · Precautionary statements

Keep away from heat. Take any precaution to avoid mixing with combustibles. Wear protective gloves / eye protection / face protection. Keep/Store away from clothing/combustible materials.

(Contd. on page 2)

CA

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

(Contd. of page 1)

Trade name: Callisto Implant 78

In case of fire: Use for extinction: CO2, powder or water spray.

Dispose of contents/container in accordance with local/regional/national/international regulations. · Additional information: Alloys do not require a label, providing they do not present a hazard to human health by inhalation, ingestion or contact with skin or to the aquatic environment in the form in which they are placed on the market. · Classification system: · NFPA ratings (scale 0 - 4) Health = 0Fire = 0*Reactivity* = 0· HMIS-ratings (scale 0 - 4) Health = 0Fire = 0**REACTIVITY O** *Reactivity* = 0· Results of PBT and vPvB assessment

· **PBT:** Not applicable.

0

0

HEALTH

· Other hazards

FIRE

· vPvB: Not applicable.

3 Composition/information on ingredients

- · Chemical characterization: Mixtures
- · Description: Dental alloy

· Dangerous components:			
7440-06-4	platinum	5-<10%	
7440-05-3	palladium	5-<10%	
7440-74-6	indium	3-5%	

4 First-aid measures

- · Description of first aid measures
- · General information: No special measures required.
- · After inhalation:

Grinding dust: Supply fresh air; consult doctor in case of complaints.

- In case of unconsciousness place patient stably in side position for transportation.
- After skin contact: After contact with the molten product, cool rapidly with cold water.
- · After eye contact:

Grinding dust: Rinse opened eye for several minutes under running water. If symptoms persist, consult a doctor.

- Mechanical effects only.
- · After swallowing: Seek medical treatment.
- · Information for doctor:
- · Most important symptoms and effects, both acute and delayed No further relevant information available.
- · Indication of any immediate medical attention and special treatment needed
- No further relevant information available.

(Contd. on page 3)

CA

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 78

(Contd. of page 2)

5 Fire-fighting measures

- · Extinguishing media
- · Suitable extinguishing agents: The product is not flammable.
- Special hazards arising from the substance or mixture No further relevant information available.
- Advice for firefighters
- · Protective equipment: Do not inhale explosion gases or combustion gases.

6 Accidental release measures

- · Personal precautions, protective equipment and emergency procedures Not required.
- · Environmental precautions: No special measures required.
- · Methods and material for containment and cleaning up: Pick up mechanically.
- · Reference to other sections
- See Section 7 for information on safe handling.
- See Section 8 for information on personal protection equipment.
- See Section 13 for disposal information.

7 Handling and storage

- · Handling:
- · Precautions for safe handling
- *Extractors are required on all machines used for thermal processing or splinter removal processes. Information about protection against explosions and fires:* No special measures required.
- · Conditions for safe storage, including any incompatibilities
- · Storage:
- Requirements to be met by storerooms and receptacles: No special requirements.
- · Information about storage in one common storage facility: Not required.
- · Further information about storage conditions: None.
- Specific end use(s) No further relevant information available.

8 Exposure controls/personal protection

- Additional information about design of technical systems: No further data; see item 7.
- · Control parameters
- · Components with limit values that require monitoring at the workplace:

7440-06-4 platinum

- *EL* Long-term value: 1* 0.002** mg/m³ *metal;**soluble salts (as Pt): S
- *EV* Long-term value: 1* 0.002** mg/m³ *metal;**water-soluble compounds (as Pt)

7440-74-6 indium

- *EL* Long-term value: 0.1 mg/m³ as In; ACGIH 2B
- *EV* Long-term value: 0.1 mg/m³ as indium
- · Additional information: The lists that were valid during the creation were used as basis.

(Contd. on page 4)

CA

Reviewed on 05/07/2015

Printing date 06/23/2015

Version number 4

Trade name: Callisto Implant 78

(Contd. of page 3)

· Exposure controls

· Personal protective equipment:

• General protective and hygienic measures: Usual hygienic measures for dental practice.

- Breathing equipment: Use respiratory protective device against the effects of fumes/dust/aerosol.
- Protection of hands:
- Protective gloves should always be worn during mechanical and thermal processing. After use of gloves apply skin-cleaning agents and skin cosmetics.
- Material of gloves Mechanical processing: Leather gloves Strong gloves Thermal processing:
- Heat protection gloves
- · Penetration time of glove material

The exact break through time has to be found out by the manufacturer of the protective gloves and has to be observed.

· Eye protection:

Always wear safety goggles during mechanical processing (grinding, sawing /cutting, drilling, milling).

Information on basic physical and	chemical properties	
General Information		
Appearance:		
Form:	Solid White	
Color: Odor:	White Odorless	
Odour: Odour threshold:	Not determined.	
pH-value:	Not applicable.	
Change in condition		
Melting point/Melting range:	1135–1250 °C	
Boiling point/Boiling range:	Undetermined.	
Flash point:	Not applicable.	
Flammability (solid, gaseous):	Not determined.	
Auto igniting:	Product is not selfigniting.	
Danger of explosion:	Product does not present an explosion hazard.	
Explosion limits:		
Lower:	Not determined.	
Upper:	Not determined.	
Vapor pressure:	Not applicable.	
Density at 20 °C:	17.6 g/cm ³	
Relative density	Not determined.	
Vapour density	Not applicable.	
Evaporation rate	Not applicable.	
Solubility in / Miscibility with		
Water:	Insoluble.	

- ć

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 78

		(Contd. of page
· Viscosity:		
Dynamic:	Not applicable.	
Kinematic:	Not applicable.	
· Solvent content:		
Solids content:	100 %	
• Other information	No further relevant information available.	

10 Stability and reactivity

 $\cdot \textit{Reactivity} \ \textit{No further relevant information available}.$

- *Chemical stability Stable under normal handling and storage conditions.*
- Thermal decomposition / conditions to be avoided: No decomposition if used according to specifications.
- · Possibility of hazardous reactions No dangerous reactions known.
- · Conditions to avoid No further relevant information available.
- $\cdot \textit{Incompatible materials: } No further relevant information available.$
- · Hazardous decomposition products: None under normal conditions of storage and use.

11 Toxicological information

- · Information on toxicological effects
- · Acute toxicity:
- Primary irritant effect:
- on the skin: No irritant effect.
- \cdot on the eye: No irritating effect.
- Sensitization: No sensitizing effects known.
- Additional toxicological information:

Fumes or dusts generated from cutting or grinding operations may cause respiratory irritation.

· Carcinogenic categories

· NTP (National Toxicology Program)

None of the ingredients is listed.

· OSHA-Ca (Occupational Safety & Health Administration)

None of the ingredients is listed.

12 Ecological information

- · Toxicity
- Aquatic toxicity: No further relevant information available.
- · Persistence and degradability No further relevant information available.
- · Behavior in environmental systems:
- · Bioaccumulative potential No further relevant information available.
- · Mobility in soil No further relevant information available.
- · Additional ecological information:
- · General notes: Generally not hazardous for water
- · Results of PBT and vPvB assessment
- · **PBT:** Not applicable.
- **vPvB:** Not applicable.

(Contd. on page 6)

Version number 4

Reviewed on 05/07/2015

(Contd. of page 5)

Trade name: Callisto Implant 78

• Other adverse effects No further relevant information available.

13 Disposal considerations

· Waste treatment methods

· Recommendation:

Printing date 06/23/2015

Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.

· Uncleaned packagings:

• Recommendation: Disposal must be made according to official regulations.

· UN-Number	
· DOT	NA3077
· TDG, ADN, IMDG, IATA	Void
· UN proper shipping name	
· TDG, ADN, IMDG, IATA	Void
· Transport hazard class(es)	
DOT	
Class	9 Miscellaneous dangerous substances and articles
Label	9
TDG, ADN, IMDG, IATA	
Class	Void
Packing group	
DOT	III
TDG, IMDG, IATA	Void
Environmental hazards:	
Marine pollutant:	No
Special precautions for user	Not applicable.
Transport in bulk according to Anne:	x II of
MARPOL73/78 and the IBC Code	Not applicable.
Transport/Additional information:	Product is not classified as a dangerous good for transpor (ADR, IMDG, IATA).

15 Regulatory information

· Safety, health and environmental regulations/legislation specific for the substance or mixture

· Sara

· Section 355 (extremely hazardous substances):

None of the ingredients is listed.

• Section 313 (Specific toxic chemical listings):

7440-22-4 silver

7440-50-8 copper

(Contd. on page 7)

CA

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 78

(Contd. of page 6)

```	cic Substances	Control Act)
7440-57-5	Gold	

 7440-22-4
 silver

 7440-50-8
 copper

 7440-06-4
 platinum

7440-74-6 indium 7439-88-5 iridium

· Proposition 65

· Chemicals known to cause cancer:

None of the ingredients is listed.

 $\cdot$  Chemicals known to cause reproductive toxicity for females:

None of the ingredients is listed.

· Chemicals known to cause reproductive toxicity for males:

None of the ingredients is listed.

· Chemicals known to cause developmental toxicity:

None of the ingredients is listed.

#### · Carcinogenic categories

· EPA (Environmental Protection Agency)

 7440-22-4
 silver
 D

 7440-50-8
 copper
 D

 • TLV (Threshold Limit Value established by ACGIH)
 D

 None of the ingredients is listed.
 .

 • NIOSH-Ca (National Institute for Occupational Safety and Health)
 .

None of the ingredients is listed.

#### · GHS label elements

The product is classified and labeled according to the Globally Harmonized System (GHS).

· Hazard pictograms



- · Signal word Danger
- · Hazard statements

May intensify fire; oxidizer.

· Precautionary statements

Keep away from heat.

Take any precaution to avoid mixing with combustibles.

Wear protective gloves / eye protection / face protection.

Keep/Store away from clothing/combustible materials.

In case of fire: Use for extinction: CO2, powder or water spray.

Dispose of contents/container in accordance with local/regional/national/international regulations.

#### · National regulations:

• Other regulations, limitations and prohibitive regulations The product is a medical device according to the Directive 93/42/EEC. This product is classified as a medical device under US and Canadian regulations and has been reviewed by the US Food and Drug Administration and Health Canada.

(Contd. on page 8)

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 78

(Contd. of page 7)

· Chemical safety assessment: A Chemical Safety Assessment has not been carried out.

#### **16 Other information**

This information is based on our present knowledge. However, this shall not constitute a guarantee for any specific product features and shall not establish a legally valid contractual relationship.

• Date of preparation / last revision 06/23/2015 / 3 • Abbreviations and acronyms:

IMDG: International Maritime Code for Dangerous Goods
DOT: US Department of Transportation
IATA: International Air Transport Association
ACGIH: American Conference of Governmental Industrial Hygienists
EINECS: European Inventory of Existing Commercial Chemical Substances
ELINCS: European List of Notified Chemical Substances
CAS: Chemical Abstracts Service (division of the American Chemical Society)
NFPA: National Fire Protection Association (USA)
HMIS: Hazardous Materials Identification System (USA)
Ox. Sol. 2: Oxidising Solids, Hazard Category 2
* Data compared to the previous version altered.