

**Safety Data Sheet**  
according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018



**1 Identification**

- **Product identifier**
- **Trade name: Telio CS Cem Implant (Base)**
- **Application of the substance / the mixture**  
Semi-permanent, self-curing dental luting composite with light-curing option for the cementation of restorations on implant abutments
- **Details of the supplier of the safety data sheet**
- **Manufacturer/Supplier:**  
Ivoclar Vivadent Inc.  
175 Pineview Drive, Amherst, N.Y. 14228  
USA  
Tel. +1 800 533 6825  
Fax +1 716 691 2285
  
- Ivoclar Vivadent Inc.  
1-6600 Dixie Road  
Mississauga, Ontario  
L5T 2Y2  
Canada  
Phone: +1 905 670 8499  
Fax: +1 905 670 3102
- **Information department: Quality Assurance / Regulatory Affairs**
- **Emergency telephone number:**  
24 Hour Emergency Assistance:  
Emergency-Call USA - Infotrac: 1-800-535-5053  
Emergency-Call Canada - Canutec: 1-613-996-6666
  
- General SDS Assistance:  
US: 1-800-533-6825  
Canada: 1-800-263-8182

**2 Hazard identification**

- **Classification of the substance or mixture**  
Skin Irritation - Category 2 H315 Causes skin irritation.  
Eye Irritation - Category 2A H319 Causes serious eye irritation.  
Skin Sensitizer - Category 1 H317 May cause an allergic skin reaction.  
Carcinogenicity – Category 2 H351 Suspected of causing cancer.

---

  - **Label elements**
  - **GHS label elements**  
The product is classified and labeled according to the Globally Harmonized System (GHS).
  - **Hazard pictograms**
- 


GHS07    GHS08
- **Signal word** Warning
  - **Hazard-determining components of labeling:**  
1,10-decandiol dimethacrylate  
titanium dioxide

(Contd. on page 2)

## Safety Data Sheet

### according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 1)

urethane dimethacrylate

· **Hazard statements**

Causes skin irritation.

Causes serious eye irritation.

May cause an allergic skin reaction.

Suspected of causing cancer.

· **Precautionary statements**

Obtain special instructions before use.

Do not handle until all safety precautions have been read and understood.

Avoid breathing dust/fume/gas/mist/vapours/spray.

Wash thoroughly after handling.

Contaminated work clothing should not be allowed out of the workplace.

Wear protective gloves/protective clothing/eye protection/face protection.

If on skin: Wash with plenty of water.

Specific treatment (see on this label).

**IF IN EYES:** Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

**IF exposed or concerned:** Get medical advice/attention.

Take off contaminated clothing and wash it before reuse.

If skin irritation or rash occurs: Get medical advice/attention.

If eye irritation persists: Get medical advice/attention.

Store locked up.

Dispose of contents/container in accordance with local/regional/national/international regulations.

· **Additional information:**

Medical devices are exempt from the labeling provisions of GHS according to the Hazard Communication Standard (HCS).

· **Classification system:**

· **NFPA ratings (scale 0 - 4)**



Health = 2

Fire = 0

Reactivity = 0

· **HMIS-ratings (scale 0 - 4)**



HEALTH 2 Health = 2

FIRE 0 Fire = 0

REACTIVITY 0 Reactivity = 0

### 3 Composition/Information on ingredients

· **Chemical characterization: Mixtures**

· **Description:** Mixture of the substances listed below with nonhazardous additions.

· **Dangerous components:**

CAS: 13760-80-0	ytterbium trifluoride	10-<20% w/w *
CAS: 6701-13-9	1,10-decandiol dimethacrylate	10-<20% w/w *
CAS: 72869-86-4	urethane dimethacrylate	1-1.5% w/w *
CAS: 13463-67-7	titanium dioxide	1-1.5% w/w *
CAS: 3380-34-5	triclosan	0.3-<1% w/w *

\* Actual concentration ranges are withheld as a trade secret.

CA

(Contd. on page 3)

## Safety Data Sheet according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 2)

### 4 First aid measures

- **Description of first aid measures**
- **After inhalation:**  
Supply fresh air and to be sure call for a doctor.  
In case of unconsciousness place patient stably in side position for transportation.
- **After skin contact:** Rinse with water.
- **After eye contact:**  
Rinse opened eye for several minutes under running water. If symptoms persist, consult a doctor.
- **After swallowing:** If symptoms persist consult doctor.
- **Information for doctor:**
- **Most important symptoms and effects, both acute and delayed** No further relevant information available.
- **Indication of any immediate medical attention and special treatment needed**  
No further relevant information available.

### 5 Firefighting measures

- **Extinguishing media**
- **Suitable extinguishing agents:**  
CO<sub>2</sub>, extinguishing powder or water spray. Fight larger fires with water spray or alcohol resistant foam.
- **Special hazards arising from the substance or mixture** No further relevant information available.
- **Advice for firefighters**
- **Protective equipment:** No special measures required.

### 6 Accidental release measures

- **Personal precautions, protective equipment and emergency procedures** Not required.
- **Environmental precautions:** Do not allow to enter sewers/ surface or ground water.
- **Methods and material for containment and cleaning up:** Pick up mechanically.
- **Reference to other sections**  
See Section 7 for information on safe handling.  
See Section 8 for information on personal protection equipment.  
See Section 13 for disposal information.

### 7 Handling and storage

- **Handling:**
- **Precautions for safe handling**  
Only adequately trained personnel should handle this product.  
For use in dentistry only.
- **Information about protection against explosions and fires:** No special measures required.
- **Conditions for safe storage, including any incompatibilities**
- **Storage:**
- **Requirements to be met by storerooms and receptacles:** Store only in the original receptacle.
- **Information about storage in one common storage facility:** Not required.
- **Further information about storage conditions:**  
Store in a cool place.  
Protect from exposure to the light.  
Protect from heat and direct sunlight.  
Keep receptacle tightly sealed.

(Contd. on page 4)

CA

## Safety Data Sheet according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 3)

· **Specific end use(s)** No further relevant information available.

### 8 Exposure controls/ Personal protection

- **Additional information about design of technical systems:** No further data; see item 7.
- **Control parameters**
- **Components with limit values that require monitoring at the workplace:**  
The product does not contain any relevant quantities of materials with critical values that have to be monitored at the workplace.
- **Additional information:** The lists that were valid during the creation were used as basis.
- **Exposure controls**
- **Personal protective equipment:**
- **General protective and hygienic measures:**  
Usual hygienic measures for dental practice and dental laboratories.  
Wash hands before breaks and at the end of work.  
Keep away from foodstuffs, beverages and feed.  
Avoid contact with the skin.
- **Breathing equipment:** Not required.
- **Protection of hands:**



Protective gloves

The glove material has to be impermeable and resistant to the product/ the substance/ the preparation.  
Due to missing tests no recommendation to the glove material can be given for the product/ the preparation/ the chemical mixture.

Selection of the glove material on consideration of the penetration times, rates of diffusion and the degradation

- **Material of gloves**  
The selection of the suitable gloves does not only depend on the material, but also on further marks of quality and varies from manufacturer to manufacturer. As the product is a preparation of several substances, the resistance of the glove material can not be calculated in advance and has therefore to be checked prior to the application.
- **Penetration time of glove material**  
The exact break through time has to be found out by the manufacturer of the protective gloves and has to be observed.
- **Eye protection:** Safety glasses

### 9 Physical and chemical properties

- **Information on basic physical and chemical properties**
- **General Information**
- **Appearance:**

Form:	Pasty
Color:	According to product specification
- **Odor:** Characteristic
- **Odor threshold:** Not determined.
- **pH-value:** Not determined.
- **Change in condition**

Melting point/Melting range:	Undetermined.
------------------------------	---------------

(Contd. on page 5)

CA

## Safety Data Sheet

### according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 4)

<b>Boiling point/Boiling range:</b>	Undetermined.
· <b>Flash point:</b>	Not applicable.
· <b>Flammability (solid, gaseous):</b>	Not applicable.
· <b>Auto igniting:</b>	Product is not selfigniting.
· <b>Danger of explosion:</b>	Product does not present an explosion hazard.
· <b>Explosion limits:</b>	
<b>Lower:</b>	Not determined.
<b>Upper:</b>	Not determined.
· <b>Vapor pressure:</b>	Not determined.
· <b>Density at 20 °C:</b>	1.47 g/cm <sup>3</sup>
· <b>Relative density</b>	Not determined.
· <b>Vapor density</b>	Not determined.
· <b>Evaporation rate</b>	Not determined.
· <b>Solubility in / Miscibility with Water:</b>	Nearly insoluble.
· <b>Partition coefficient (n-octanol/water):</b>	Not determined.
· <b>Viscosity:</b>	
<b>Dynamic:</b>	Not determined.
<b>Kinematic:</b>	Not determined.
· <b>Other information</b>	No further relevant information available.

### 10 Stability and reactivity

- **Reactivity** No further relevant information available.
- **Chemical stability**  
Stable under normal handling and storage conditions.  
Photoreactive (Danger of polymerization).
- **Thermal decomposition / conditions to be avoided:** No decomposition if used according to specifications.
- **Possibility of hazardous reactions** No dangerous reactions known.
- **Conditions to avoid**  
Protect from exposure to the light.  
Keep away from heat and direct sunlight.
- **Incompatible materials:** No further relevant information available.
- **Hazardous decomposition products:** None under normal conditions of storage and use.

### 11 Toxicological information

- **Information on toxicological effects**
- **Acute toxicity:**
- **on the skin:** No irritant effect.
- **on the eye:** No irritating effect.
- **Sensitization:** Sensitization possible through skin contact.
- **Additional toxicological information:** No further relevant information available.
- **Carcinogenic categories**

· **IARC (International Agency for Research on Cancer)**

CAS: 7631-86-9 | silicon dioxide, chemically prepared

3

(Contd. on page 6)

CA

## Safety Data Sheet

### according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 5)

CAS: 13463-67-7	titanium dioxide	2B
CAS: 128-37-0	Butylated hydroxytoluene	3
· <b>NTP (National Toxicology Program)</b>		
None of the ingredients is listed.		

### 12 Ecological information

- **Toxicity**
- **Aquatic toxicity:** No further relevant information available.
- **Persistence and degradability** No further relevant information available.
- **Behavior in environmental systems:**
- **Bioaccumulative potential** No further relevant information available.
- **Mobility in soil** No further relevant information available.
- **Ecotoxicological effects:**
- **Remark:** Very toxic for fish
- **Additional ecological information:**
- **General notes:**  
*Water hazard class 1 (Self-assessment): slightly hazardous for water*  
*Do not allow undiluted product or large quantities of it to reach ground water, water course or sewage system.*  
*Also poisonous for fish and plankton in water bodies.*  
*Very toxic for aquatic organisms*
- **Results of PBT and vPvB assessment**
- **PBT:** Not applicable.
- **vPvB:** Not applicable.
- **Other adverse effects** No further relevant information available.

### 13 Disposal considerations

- **Waste treatment methods**
- **Recommendation:**  
*Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.*
- **Uncleaned packagings:**
- **Recommendation:** Disposal must be made according to official regulations.

### 14 Transport information

- **UN-Number**
- **DOT/TDG, ADR, ADN, IMDG, IATA**      *Void*
- **UN proper shipping name**
- **DOT/TDG, ADR, ADN, IMDG, IATA**      *Void*
- **Transport hazard class(es)**
- **DOT, ADR, ADN, IMDG, IATA**
- **Class**      *Void*
- **Packing group**
- **DOT/TDG, ADR, IMDG, IATA**      *Void*

(Contd. on page 7)

CA

## Safety Data Sheet

according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 6)

· <b>Environmental hazards:</b>	
· <b>Marine pollutant:</b>	No
· <b>Special precautions for user</b>	Not applicable.
· <b>Transport in bulk according to Annex II of MARPOL73/78 and the IBC Code</b>	Not applicable.
· <b>Transport/Additional information:</b>	Product is not classified as a dangerous good for transport (ADR, IMDG, IATA).
· <b>UN "Model Regulation":</b>	Void

### 15 Regulatory information

· **Safety, health and environmental regulations/legislation specific for the substance or mixture**

· **Sara**

· **Section 355 (extremely hazardous substances):**

None of the ingredients is listed.

· **Section 313 (Specific toxic chemical listings):**

None of the ingredients is listed.

· **TSCA (Toxic Substances Control Act):**

CAS: 13760-80-0 ytterbium trifluoride

CAS: 6701-13-9 1,10-decandiol dimethacrylate

CAS: 7631-86-9 silicon dioxide, chemically prepared

CAS: 72869-86-4 urethane dimethacrylate

CAS: 13463-67-7 titanium dioxide

CAS: 10287-53-3 ethyl p-dimethylaminobenzoate

CAS: 3380-34-5 triclosan

CAS: 41637-38-1 Bis-EMA

CAS: 10373-78-1 camphorquinone

CAS: 128-37-0 Butylated hydroxytoluene

· **GHS label elements**

The product is classified and labeled according to the Globally Harmonized System (GHS).

· **Hazard pictograms**



GHS07 GHS08

· **Signal word** Warning

· **Hazard-determining components of labeling:**

1,10-decandiol dimethacrylate

titanium dioxide

urethane dimethacrylate

· **Hazard statements**

Causes skin irritation.

Causes serious eye irritation.

May cause an allergic skin reaction.

(Contd. on page 8)

CA



## Safety Data Sheet

### according to HPR, Schedule 1

Printing date 07/22/2019

Version number 2

Reviewed on 04/26/2018

**Trade name: Telio CS Cem Implant (Base)**

(Contd. of page 7)

*Suspected of causing cancer.*

· **Precautionary statements**

*Obtain special instructions before use.*

*Do not handle until all safety precautions have been read and understood.*

*Avoid breathing dust/fume/gas/mist/vapours/spray.*

*Wash thoroughly after handling.*

*Contaminated work clothing should not be allowed out of the workplace.*

*Wear protective gloves/protective clothing/eye protection/face protection.*

*If on skin: Wash with plenty of water.*

*Specific treatment (see on this label).*

*IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.*

*IF exposed or concerned: Get medical advice/attention.*

*Take off contaminated clothing and wash it before reuse.*

*If skin irritation or rash occurs: Get medical advice/attention.*

*If eye irritation persists: Get medical advice/attention.*

*Store locked up.*

*Dispose of contents/container in accordance with local/regional/national/international regulations.*

· **National regulations:**

· **Other regulations, limitations and prohibitive regulations**

*The product is a medical device according to the Directive 93/42/EEC.*

*This product is classified as a medical device under US and Canadian regulations and has been reviewed by the US Food and Drug Administration and Health Canada.*

· **Chemical safety assessment:** *A Chemical Safety Assessment has not been carried out.*

### 16 Other information

*This information is based on our present knowledge. However, this shall not constitute a guarantee for any specific product features and shall not establish a legally valid contractual relationship.*

· **Date of the latest revision of the safety data sheet** 07/22/2019 / 1

· **Abbreviations and acronyms:**

*IMDG: International Maritime Code for Dangerous Goods*

*DOT: US Department of Transportation*

*IATA: International Air Transport Association*

*EINECS: European Inventory of Existing Commercial Chemical Substances*

*ELINCS: European List of Notified Chemical Substances*

*CAS: Chemical Abstracts Service (division of the American Chemical Society)*

*NFPA: National Fire Protection Association (USA)*

*HMIS: Hazardous Materials Identification System (USA)*

*PBT: Persistent, Bioaccumulative and Toxic*

*vPvB: very Persistent and very Bioaccumulative*

· **\* Data compared to the previous version altered.**

CA