



BioUniversal®

Reduced-gold universal alloy

to be veneered with special ceramic materials and composite materials.

Au 59.4	Pt 2.0	Pd 9.5	Ag 25.5	Zn 2.0	Sn <1.0	Fe <1.0	Re <1.0	Ru <1.0
------------	-----------	-----------	------------	-----------	------------	------------	------------	------------

Advantages

- Copper-free
- Yellow color
- Compatible with low fusing special ceramic and composite materials
- High temperature strength
- Excellent milling properties

Indication

Onlays, $\frac{3}{4}$ crowns, PFM Crowns, telescope and conus crowns, posts, short- and long-span bridges, implant superstructures

Technical Data

Color	yellow
Type	4
Density (g/cm ³)	14.5
Melting range (°C)	1020 – 1100
Casting temperature (°C)	1155 – 1215
CTE 25 – 500 °C	16.1
CTE 20 – 600 °C	16.3
Elongation (%)	6.0
Modulus of elasticity (MPa)	103.000
Oxide firing °C / min. / vacuum	700 / 3 / vacuum
Vickers hardness	220
0.2 % Proof stress (MPa)	480



Clinical case by Dentallabor K. Czirjak GmbH, Germany

Certificate

Test material: BioUniversal Alloys

Composition in % weight	Au	Pt	Pd	Ag	Cu	In	Ir	Zn	Other
BioUniversal® PKF	75.0	8.6	–	11.9	–	<1.0	<1.0	2.0	Fe <1.0, Ta <1.0, Rh <1.0
BioUniversal® PdF	71.1	9.2	–	11.7	4.5	1.3	<1.0	1.5	Fe <1.0, Ta <1.0
BioUniversal®	59.4	2.0	9.5	25.5	–	–	–	2.0	Fe <1.0, Re <1.0, Ru <1.0, Sn <1.0
BioUniversal® E	–	–	40.0	52.2	–	1.4	–	<1.0	Sn <5.5, Re <1.0, Ru <1.0

Manufacturer

Ivoclar Vivadent Inc., 175 Pineview Drive, Amherst, NY 14228, USA

Corrosion resistance

The test was conducted according to the international regulations of ISO 1562 and ISO 6871-1: static immersion test through analytical determination of the metal ion release after a 7-day immersion.

Test results: The metal ion release after 7 days of immersion was not significant.

Testing facility: Louisiana State University, Dr. Sakar

Cytotoxicity

The Agar Diffusion test determines the biological reactivity of cell culture on test material.

Test results: The test material is considered non-cytotoxic and meets the requirements of the Agar Diffusion test according to ISO 10993-5.

Mutagenicity

An Ames assay was conducted to determine any possible cancer potential.

Test results: No mutagenicity potential was found to exist in these BioUniversal alloys.

Kligman Maximization

This test evaluated the allergenic potential and/or sensitizing capacity of these BioUniversal alloys.

Test results: Based on the standards set by the study protocol, these alloys exhibited no reaction to the challenge (0 % sensitization).

Sensitivity of oral mucosa

Test to determine the contact sensitivity of these BioUniversal alloys at the buccal oral mucosa.

Test results: No reactions were noted in conjunction with these BioUniversal alloys.

Testing facility: Toxikon Corporation, 15 Wiggins Avenue, Bedford, Massachusetts

Amherst, May 2010



Dr. George Tysowsky, D. D. S., M. P. H.
Vice President-Technology

EC Representative
Ivoclar Vivadent AG
Bendererstrasse 2
FL-9494 Schaan
Fürstentum Liechtenstein
Tel. +423 / 235 35 35
Fax +423 / 235 33 60
www.ivoclarvivadent.com

Manufacturer:
Ivoclar Vivadent, Inc.
175 Pineview Drive
USA, Amherst, NY 14228
Tel. (716) 691 0010
Tel. 1 800 533 6825
Fax (716) 691 2285

Notified Body:
SGS United Kingdom Limited
Unit 202B
Worle Parkway
Weston Super Mare BS22 6WA
United Kingdom



passion vision innovation