

# Combinations by Indication

	Esthetic Luting Composite System		Universal Luting Composite System	Self-adhesive Resin Cement
Cement	Variolink® Veneer	Variolink® II	Multilink® Automix	SpeedCEM
Curing Mode	Light-cure	Light/Dual-cure	Self-cure with light-curing option	Self-cure with light-curing option
Cementation Method	Adhesive: Syntac or Excite F Total-Etch	Adhesive: Syntac or Excite F DSC Total-Etch	Adhesive: Multilink Primer A/B Self-Etch	Self-Adhesive
<b>Metal/Metal-based</b>				
Inlays/Onlays	-	-	✓	✓
Crowns/Bridges	-	-	✓	✓
Endodontic Posts	-	-	✓	✓
Maryland Bridges	-	-	✓	-
<b>Glass-Ceramics (eg leucite reinforced)</b>				
Inlays/Onlays	✓	✓	✓	-
Veneers	✓	✓	-	-
Crowns	-	✓	✓	-
<b>Lithium disilicate (LS<sub>2</sub>)</b>				
Occlusal Veneers	✓	✓	-	-
Thin Veneers/Veneers	✓	✓	-	-
Inlays/Onlays	✓	✓	✓	-
Partial Crowns	-	✓	✓	-
Crowns	-	✓	✓	✓
3-unit Bridges	-	✓	✓	✓
<b>Oxide-Ceramics (Zirconia / Alumina)</b>				
Crowns/Bridges	-	-	✓	✓
Endodontic Posts	-	-	✓	✓
<b>Indirect Composites</b>				
Inlays/Onlays	✓	✓	✓	-
Crowns	-	✓	✓	-
Endodontic Posts	-	✓	✓	✓



Recommended product combination

- Not recommended

# Combinations by Products

	Esthetic Luting Composite System		Universal Luting Composite System	Self-adhesive Resin Cement
Cement	Variolink® Veneer	Variolink® II	Multilink® Automix	SpeedCEM®
Curing Mode	Light-cure	Light/Dual-cure	Self-cure with light-curing option	Self-cure with light-curing option
Cementation Method	Adhesive: Syntac® or ExciTE® F Total-Etch	Adhesive: Syntac® or ExciTE® F DSC Total-Etch	Adhesive: Multilink® Primer A/B Self-Etch	Self-Adhesive
IPS e.max® Press	✓	✓	✓	✓*
IPS e.max® CAD	✓	✓	✓	✓*
IPS e.max® Ceram Veneers	✓	✓	–	–
IPS e.max® ZirCAD	–	–	✓	✓
IPS Empress® Esthetic	✓	✓	✓	–
IPS Empress® CAD	✓	✓	✓	–
SR Nexco® (metal-based)	–	–	✓	✓
SR Nexco® (metal-free)	✓	✓	✓	–
IPS d.SIGN®				
IPS InLine® / IPS InLine PoM®	–	–	✓	✓
IPS Classic®				
FRC Postec® Plus	–	✓	✓	✓
CosmoPost®	–	–	✓	✓



Please take note of the corresponding Instructions for Use.

- ✓ Recommended product combination
- Recommended depending on indication
- Not recommended