

Safety Data Sheet
acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

1 Identification


- **Product identifier**
- **Trade name:** ***Callisto Implant 60***
- **Relevant identified uses of the substance or mixture and uses advised against**
No further relevant information available.
- **Application of the substance / the mixture** Manufacture of dental prothesis
- **Details of the supplier of the safety data sheet**
- **Manufacturer/Supplier:**
Ivoclar Vivadent Inc.
175 Pineview Drive, Amherst, N.Y. 14228
USA
Tel. +1 800 533 6825
Fax +1 716 691 2285

- Ivoclar Vivadent Inc.
1-6600 Dixie Road
Mississauga, Ontario
L5T 2Y2
Canada
Phone: +1 905 670 8499
Fax: +1 905 670 3102
- **Information department:** Quality Assurance / Regulatory Affairs
- **Emergency telephone number:**
24 Hour Emergency Assistance:
Emergency-Call USA - Infotrac: 1-800-535-5053
Emergency-Call Canada - Canutec: 1-613-996-6666


- General SDS Assistance:
US: 1-800-533-6825
Canada: 1-800-263-8182

2 Hazard(s) identification

- **Classification of the substance or mixture**
Ox. Sol. 2 H272 May intensify fire; oxidizer.
Skin Irrit. 2 H315 Causes skin irritation.
Eye Irrit. 2A H319 Causes serious eye irritation.

 - **Label elements**
 - **GHS label elements**
The product is classified and labeled according to the Globally Harmonized System (GHS).
 - **Hazard pictograms**
- 

GHS03



GHS07
- **Signal word** Danger
 - **Hazard statements**
May intensify fire; oxidizer.
Causes skin irritation.
Causes serious eye irritation.

(Contd. on page 2)

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: **Callisto Implant 60**

(Contd. of page 1)

· **Precautionary statements**

Keep away from heat.

Take any precaution to avoid mixing with combustibles.

Wear protective gloves / eye protection / face protection.

Wear protective gloves.

Wear eye protection / face protection.

Keep/Store away from clothing/combustible materials.

Wash thoroughly after handling.

If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do.

Continue rinsing.

Specific treatment (see on this label).

If skin irritation occurs: Get medical advice/attention.

If eye irritation persists: Get medical advice/attention.

In case of fire: Use for extinction: CO₂, powder or water spray.

If on skin: Wash with plenty of water.

Take off contaminated clothing and wash it before reuse.

Dispose of contents/container in accordance with local/regional/national/international regulations.

· **Additional information:**

Alloys do not require a label, providing they do not present a hazard to human health by inhalation, ingestion or contact with skin or to the aquatic environment in the form in which they are placed on the market.

· **Classification system:**

· **NFPA ratings (scale 0 - 4)**



Health = 1

Fire = 0

Reactivity = 0

· **HMIS-ratings (scale 0 - 4)**



Health = 1

Fire = 0

Reactivity = 0

· **Other hazards**

· **Results of PBT and vPvB assessment**

· **PBT:** Not applicable.

· **vPvB:** Not applicable.

3 Composition/information on ingredients

· **Chemical characterization: Mixtures**

· **Description:** Dental alloy

· **Dangerous components:**

7440-05-3	palladium	50-100%
7440-22-4	silver	25-50%
7440-31-5	tin	2.5-<10%
7440-74-6	indium	1-<2.5%
7440-55-3	gallium	1-<2.5%

4 First-aid measures

· **Description of first aid measures**

· **General information:** No special measures required.

(Contd. on page 3)

US

Safety Data Sheet acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 60

(Contd. of page 2)

- **After inhalation:**
Grinding dust: Supply fresh air; consult doctor in case of complaints.
In case of unconsciousness place patient stably in side position for transportation.
- **After skin contact:** After contact with the molten product, cool rapidly with cold water.
- **After eye contact:**
Grinding dust: Rinse opened eye for several minutes under running water. If symptoms persist, consult a doctor.
Mechanical effects only.
- **After swallowing:** Seek medical treatment.
- **Information for doctor:**
- **Most important symptoms and effects, both acute and delayed** No further relevant information available.
- **Indication of any immediate medical attention and special treatment needed**
No further relevant information available.

5 Fire-fighting measures

- **Extinguishing media**
- **Suitable extinguishing agents:** The product is not flammable.
- **Special hazards arising from the substance or mixture** No further relevant information available.
- **Advice for firefighters**
- **Protective equipment:** Do not inhale explosion gases or combustion gases.

6 Accidental release measures

- **Personal precautions, protective equipment and emergency procedures** Not required.
- **Environmental precautions:** No special measures required.
- **Methods and material for containment and cleaning up:** Pick up mechanically.
- **Reference to other sections**
See Section 7 for information on safe handling.
See Section 8 for information on personal protection equipment.
See Section 13 for disposal information.

7 Handling and storage

- **Handling:**
- **Precautions for safe handling**
Extractors are required on all machines used for thermal processing or splinter removal processes.
- **Information about protection against explosions and fires:** No special measures required.
- **Conditions for safe storage, including any incompatibilities**
- **Storage:**
- **Requirements to be met by storerooms and receptacles:** No special requirements.
- **Information about storage in one common storage facility:** Not required.
- **Further information about storage conditions:** None.
- **Specific end use(s)** No further relevant information available.

8 Exposure controls/personal protection

- **Additional information about design of technical systems:** No further data; see item 7.

(Contd. on page 4)

US

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 60

(Contd. of page 3)

· **Control parameters**

· **Components with limit values that require monitoring at the workplace:**

7440-22-4 silver

PEL Long-term value: 0.01 mg/m³

REL Long-term value: 0.01 mg/m³

TLV Long-term value: 0.1 mg/m³
metal: dust and fume

7440-31-5 tin

PEL Long-term value: 2 mg/m³
metal

REL Long-term value: 2 mg/m³

TLV Long-term value: 2 mg/m³
metal

7440-74-6 indium

REL Long-term value: 0.1 mg/m³
as In

TLV Long-term value: 0.1 mg/m³
as In

· **Additional information:** The lists that were valid during the creation were used as basis.

· **Exposure controls**

· **Personal protective equipment:**

· **General protective and hygienic measures:** Usual hygienic measures for dental practice.

· **Breathing equipment:** Use respiratory protective device against the effects of fumes/dust/aerosol.

· **Protection of hands:**

Protective gloves should always be worn during mechanical and thermal processing.

After use of gloves apply skin-cleaning agents and skin cosmetics.

· **Material of gloves**

Mechanical processing:

Leather gloves

Strong gloves

Thermal processing:

Heat protection gloves

· **Penetration time of glove material**

The exact break through time has to be found out by the manufacturer of the protective gloves and has to be observed.

· **Eye protection:**

Always wear safety goggles during mechanical processing (grinding, sawing /cutting, drilling, milling).

9 Physical and chemical properties

· **Information on basic physical and chemical properties**

· **General Information**

· **Appearance:**

Form: Solid

Color: White

· **Odor:** Odorless

· **Odour threshold:** Not determined.

· **pH-value:** Not applicable.

(Contd. on page 5)

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 60

(Contd. of page 4)

· Change in condition	
Melting point/Melting range:	1100–1260 °C (2012–2300 °F)
Boiling point/Boiling range:	Undetermined.
· Flash point:	Not applicable.
· Flammability (solid, gaseous):	Not determined.
· Auto igniting:	Product is not selfigniting.
· Danger of explosion:	Product does not present an explosion hazard.
· Explosion limits:	
Lower:	Not determined.
Upper:	Not determined.
· Vapor pressure:	Not applicable.
· Density at 20 °C (68 °F):	10.8 g/cm ³ (90.126 lbs/gal)
· Relative density	Not determined.
· Vapour density	Not applicable.
· Evaporation rate	Not applicable.
· Solubility in / Miscibility with Water:	Insoluble.
· Partition coefficient (n-octanol/water):	Not determined.
· Viscosity:	
Dynamic:	Not applicable.
Kinematic:	Not applicable.
· Solvent content:	
Solids content:	100 %
· Other information	No further relevant information available.

10 Stability and reactivity

- **Reactivity** No further relevant information available.
- **Chemical stability** Stable under normal handling and storage conditions.
- **Thermal decomposition / conditions to be avoided:** No decomposition if used according to specifications.
- **Possibility of hazardous reactions** No dangerous reactions known.
- **Conditions to avoid** No further relevant information available.
- **Incompatible materials:** No further relevant information available.
- **Hazardous decomposition products:** None under normal conditions of storage and use.

11 Toxicological information

- **Information on toxicological effects**
- **Acute toxicity:**
- **Primary irritant effect:**
- **on the skin:** No irritant effect.
- **on the eye:** No irritating effect.
- **Sensitization:** No sensitizing effects known.
- **Additional toxicological information:**
- Fumes or dusts generated from cutting or grinding operations may cause respiratory irritation.

(Contd. on page 6)

US

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: **Callisto Implant 60**

(Contd. of page 5)

· **Carcinogenic categories**

· **NTP (National Toxicology Program)**

None of the ingredients is listed.

· **OSHA-Ca (Occupational Safety & Health Administration)**

None of the ingredients is listed.

12 Ecological information

- **Toxicity**
- **Aquatic toxicity:** No further relevant information available.
- **Persistence and degradability** No further relevant information available.
- **Behavior in environmental systems:**
- **Bioaccumulative potential** No further relevant information available.
- **Mobility in soil** No further relevant information available.
- **Additional ecological information:**
- **General notes:** Generally not hazardous for water
- **Results of PBT and vPvB assessment**
- **PBT:** Not applicable.
- **vPvB:** Not applicable.
- **Other adverse effects** No further relevant information available.

13 Disposal considerations

- **Waste treatment methods**
- **Recommendation:**
Take to an approved landfill or a waste incineration plant, under conditions approved by the local authority.
- **Uncleaned packagings:**
- **Recommendation:** Disposal must be made according to official regulations.

14 Transport information

- | | |
|-------------------------------------|---|
| · UN-Number | |
| · DOT | NA3077 |
| · ADR, ADN, IMDG, IATA | Void |
| · UN proper shipping name | |
| · ADR, ADN, IMDG, IATA | Void |
| · Transport hazard class(es) | |
| · DOT | |
| · Class | 9 Miscellaneous dangerous substances and articles |
| · Label | 9 |
| · ADR, ADN, IMDG, IATA | |
| · Class | Void |
| · Packing group | |
| · DOT | III |
| · ADR, IMDG, IATA | Void |

(Contd. on page 7)

US

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: Callisto Implant 60

(Contd. of page 6)

· Environmental hazards:	
· Marine pollutant:	No
· Special precautions for user	Not applicable.
· Transport in bulk according to Annex II of MARPOL73/78 and the IBC Code	Not applicable.
· Transport/Additional information:	Product is not classified as a dangerous good for transport (ADR, IMDG, IATA).
· UN "Model Regulation":	-

15 Regulatory information

· **Safety, health and environmental regulations/legislation specific for the substance or mixture**

· **Sara**

· **Section 355 (extremely hazardous substances):**

None of the ingredients is listed.

· **Section 313 (Specific toxic chemical listings):**

7440-22-4 silver

7440-50-8 copper

· **TSCA (Toxic Substances Control Act):**

7440-57-5 Gold

7440-22-4 silver

7440-50-8 copper

7440-06-4 platinum

7440-74-6 indium

7439-88-5 iridium

· **Proposition 65**

· **Chemicals known to cause cancer:**

None of the ingredients is listed.

· **Chemicals known to cause reproductive toxicity for females:**

None of the ingredients is listed.

· **Chemicals known to cause reproductive toxicity for males:**

None of the ingredients is listed.

· **Chemicals known to cause developmental toxicity:**

None of the ingredients is listed.

· **Carcinogenic categories**

· **EPA (Environmental Protection Agency)**

7440-22-4 silver

D

7440-50-8 copper

D

· **TLV (Threshold Limit Value established by ACGIH)**

None of the ingredients is listed.

· **NIOSH-Ca (National Institute for Occupational Safety and Health)**

None of the ingredients is listed.

(Contd. on page 8)

Safety Data Sheet

acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: **Callisto Implant 60**

(Contd. of page 7)

· **GHS label elements**

The product is classified and labeled according to the Globally Harmonized System (GHS).

· **Hazard pictograms**



GHS03 GHS07

· **Signal word** *Danger*

· **Hazard statements**

May intensify fire; oxidizer.

Causes skin irritation.

Causes serious eye irritation.

· **Precautionary statements**

Keep away from heat.

Take any precaution to avoid mixing with combustibles.

Wear protective gloves / eye protection / face protection.

Wear protective gloves.

Wear eye protection / face protection.

Keep/Store away from clothing/combustible materials.

Wash thoroughly after handling.

If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do.

Continue rinsing.

Specific treatment (see on this label).

If skin irritation occurs: Get medical advice/attention.

If eye irritation persists: Get medical advice/attention.

In case of fire: Use for extinction: CO₂, powder or water spray.

If on skin: Wash with plenty of water.

Take off contaminated clothing and wash it before reuse.

Dispose of contents/container in accordance with local/regional/national/international regulations.

· **National regulations:**

· **Other regulations, limitations and prohibitive regulations**

The product is a medical device according to the Directive 93/42/EEC.

This product is classified as a medical device under US and Canadian regulations and has been reviewed by the US Food and Drug Administration and Health Canada.

· **Chemical safety assessment:** A Chemical Safety Assessment has not been carried out.

16 Other information

This information is based on our present knowledge. However, this shall not constitute a guarantee for any specific product features and shall not establish a legally valid contractual relationship.

· **Date of preparation / last revision** 06/23/2015 / 3

· **Abbreviations and acronyms:**

ADR: Accord européen sur le transport des marchandises dangereuses par Route (European Agreement concerning the International Carriage of Dangerous Goods by Road)

IMDG: International Maritime Code for Dangerous Goods

DOT: US Department of Transportation

IATA: International Air Transport Association

ACGIH: American Conference of Governmental Industrial Hygienists

EINECS: European Inventory of Existing Commercial Chemical Substances

ELINCS: European List of Notified Chemical Substances

CAS: Chemical Abstracts Service (division of the American Chemical Society)

NFPA: National Fire Protection Association (USA)

HMS: Hazardous Materials Identification System (USA)

Ox. Sol. 2: Oxidising Solids, Hazard Category 2

Skin Irrit. 2: Skin corrosion/irritation, Hazard Category 2

(Contd. on page 9)

Safety Data Sheet
acc. to OSHA HCS

Printing date 06/23/2015

Version number 4

Reviewed on 05/07/2015

Trade name: **Callisto Implant 60**

Eye Irrit. 2A: Serious eye damage/eye irritation, Hazard Category 2A

· *** Data compared to the previous version altered.**

(Contd. of page 8)

US